

Product Name :
Micro Plate Reader**Product Code :**
CLNCL-CDD0003**Description :**

Micro Plate Reader

Technical Specification :

Wavelength range (nm): 400-800;

Light source lamp: 12V/20W quartz halogen lamp (life ? 3000h), and has a sleep function;

Detection range (A): 0.000 ~ 4.000; detection track: 8 channels;

Filter configuration (nm): 4 pieces of standard configuration: 405, 450, 492, 630, up to 10 filters can be selected in the range of 400-800;

Reading speed: 5 seconds / 96 holes (single wavelength); 10 seconds / 96 holes (dual wavelength);

Wavelength characteristics: The accuracy of the center wavelength of the filter configured by the analyzer should not exceed $\pm 2\text{nm}$; the half width of the wavelength (nm): 7 ± 2 ;Absorbance accuracy (A): ± 0.005 (when the absorbance ranges from 0.000 to ≈ 0.500);

Linearity error: linear correlation coefficient (r) ? 0.995 (in the range of 0 to 3.000 absorbance value);

Absorbance repeatability of the instrument: CV ? 0.5%;

Stability of absorbance of the instrument (A): ? ± 0.005 ;

Absorbance resolution (A): 0.001;

Difference between channels: ≈ 0.02 (with air as a reference, the difference in absorbance between channels of the measuring instrument);

Display: 5.6 inch color touch screen;

Operation interface: with power-on self-test function;

Operation mode: The instrument adopts touch screen operation mode, and can input at the same time;

Vibration plate function: The instrument has a vibration plate function;

Detection method: The instrument has two detection modes: single wavelength and dual wavelength, single hole and double hole;

Detection function: It has the functions of absorbance detection, qualitative detection and quantitative detection; it can visualize the functions of 12 inspection items of layout board and single board;

Detection output: qualitative: sample absorbance, S/CO value, critical value and negative positive judgment result; quantitative: sample absorbance, sample concentration value, normal reference value and detection judgment result; output is 96-well whole plate test result;

Calculation methods: straight line method, point-to-point method, linear regression method, semi-logarithmic regression method, exponential regression method, full logarithmic regression method, ratio regression method, ratio semi-log regression method, quadratic curve, Logit-Log curve;

Quality control function: With quality control function, it can output quality control data and L-J quality control chart Westgrad multi-rule judgment;

Storage function: program storage: 200 project programs and calibration parameters; inspection result storage: can store 100 board test results;

Communication function: the instrument has RS-232 communication interface and USB interface;

Light source signal monitoring function: It can monitor 8 channel light source signals in real time;

NAUGRA[®]

Medical Hospital Equipments Manufacturers

Website:Medical Hospital Equipments Manufacturers,