

Product Name :
Pathology Lab Microtome**Product Code :**
MED-PTH0042**Description :**

Pathology Lab Microtome

Technical Specification :

Mode: Automatic or Manual Modewithout intelligent sensing

Function: Sectioning, trimming, fast forward, fast backward, quick conversion; automatic retraction (except).

Retraction Range: 0~50?m (If 0 is set, retraction function is off)

5~50 (Retraction distance is optional in this range, and increment is 5?m, such as, 5, 10, 15, 25...)

Display: LED display:

sectioning thickness,

trimming thickness,

slice count,

slice speed.

Sectioning Thickness Range: 0.25~100?m

0.25~2.5?m increment 0.25?m

2.5~5?m increment 0.5?m

5.0~10?m increment 1?m

10~30?m increment 2?m

30~60?m increment 5?m

60~100?m increment 10?m

Trimming Thickness Range: 1~600?m

1~10?m increment 1?m

10~20?m increment 2?m

20~50?m increment 5?m

50~150?m increment 10?m

150~600?m: increment 50?m
Whole Layer Clearance Distance: 10~6000?m
(10, 12, 15, 20.....6000, after 15?m, the increment is 5?m)
Alarm when required slice finishes sectioning.
Specimen Horizontal Feed 28mm
Specimen Vertical Stroke 70mm
Specimen Holder Rotation Any angle within 360°
Specimen Orientation XY - 8°
Mini. Sectioning Thicknes 0.25?m
Max. Specimen Size 70*70mm
Feed Speed 1500?m/s-3500?m/s
Precision ±5%
Standard Accessory: Cassette Clamp or Paraffin Block Clamp(Cassette Clamp is acquiescent)
Power Supply: AC110/220V±10%, 50/60Hz

NAUGRA[®]

Medical Hospital Equipments Manufacturers

Website: [SITE,](#)