

Email: sales@naugramedical.com

Phone: 9896600003

Product Name: Pathology Lab Hot Air Oven **Product Code:** MED-PTH0049



Description:

Pathology Lab Hot Air Oven

Technical Specification:

Capacity: 30L

Temp. Range: RT+10~300°C Temp. Precision: 0.1°C Temp. Fluctuation: ±0.5°C Ambient Temp: 5~40°C Timing Range: 1~9999min

Shelves No: 2 pcs Display: LED Consumption: 800W

Power Supply: AC220V±10%, 50/60Hz;



Medical Hospital Equipments Manufacturers

Website: Medical Hospital Equipments Manufacturers,